


Auftraggeber Maulide
hazh115



Current Data Parameters
NAME 6Aug3019
EXPNO 23
PROCNO 1

F2 - Acquisition Parameters
Date_ 20190830
Time 10.15 h

INSTRUM spect
PROBHD Z126545_0016 (
PULPROG hsqcetgp
TD 1024
SOLVENT CDCl3
NS 2
DS 16
SWH 7812.500 Hz
FIDRES 15.258789 Hz
AQ 0.0655360 sec
RG 2050
DW 64.000 usec
DE 12.00 usec
TE 298.1 K
CNST2 145.0000000
D0 0.00000300 sec
D1 1.50000000 sec
D4 0.00172414 sec
D11 0.03000000 sec
D16 0.00020000 sec
IN0 0.00002010 sec
TDav 1
ZGOPTNS
SFO1 600.2537065 MHz
NUC1 1H
P1 12.25 usec
P2 24.50 usec
P28 0 usec
PLW1 25.60000038 W
SFO2 150.9443025 MHz
NUC2 13C
CPDPRG[2] garp
P3 10.00 usec
P4 20.00 usec
PCPD2 60.00 usec
PLW2 66.00000000 W
PLW12 1.83329999 W
GPNAM[1] SMSQ10.100
GPZ1 80.00 %
GPNAM[2] SMSQ10.100
GPZ2 20.10 %
P16 1000.00 usec

F1 - Acquisition parameters
TD 128
SFO1 150.9443 MHz
FIDRES 388.681580 Hz
SW 164.800 ppm
FnMODE Echo-Antiecho

F2 - Processing parameters
SI 1024
SF 600.2500000 MHz
WDW QSINE
SSB 2
LB 0 Hz
GB 0
PC 1.40

F1 - Processing parameters
SI 1024
MC2 echo-antiecho
SF 150.9329825 MHz
WDW QSINE
SSB 2
LB 0 Hz
GB 0

